

Certificate of Conformity
D.L.S. Electronic Systems

Company: HD Electric Company
Product Tested: Digital Voltage Indicator
Model Tested: DVI-100
Models Covered: DVI-100*, DVI-100T* (The * indicates various accessories)
Testing Date: October 16, 2015
Report Numbers: 15CA0211

A sample of the equipment as configured in the accompanying test report has been found to comply with the following standards:

IEC 61010-1: 2010 (Third Edition) Safety requirements for electrical equipment for measurement, control, and laboratory use” - Part 1: General requirements
IEC 61010-2-030: 2010 (First Edition) Particular requirements for testing and measuring circuits
EN 61010-1: 2010 (Third Edition) Safety requirements for electrical equipment for measurement, control, and laboratory use” - Part 1: General requirements
EN 61010-2-030: 2010 (First Edition) Particular requirements for testing and measuring circuits


Summary:

The Digital Voltage Indicator, Model DVI-100, met the requirements of IEC 61010-1, IEC 61010-2-30, EN 61010-1, EN 61010-2-30 “Safety requirements for electrical equipment for measurement, control, and laboratory use” - Part 1: General requirements and Part 2-030: Particular requirements for testing and measuring circuits

Furthermore, the manufacturer may declare that the product is in conformity with the above sections of EN 61326-1:2013.

Attested to:

D.L.S. Electronics



Brian J. Mattson
General Manager

2-10-2017
Date

