

Certificate of Conformity  
D.L.S. Electronic Systems

Company: HD Electric Company  
Product Tested: Digital Voltage Indicator  
Model Tested: DVI-100  
Models Covered: DVI-100\*, DVI-100T\* (The \* indicates various accessories)  
Testing Date: October 16, 2015  
Report Numbers: 21355 and 21357

*A sample of the equipment as configured in the accompanying test report has been found to comply with the following standards:*

EN 61326-1:2013  
ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE;  
EMC REQUIREMENTS - Part 1: General Requirements  
(IEC 61326-1:2012) - Industrial Locations

USING IEC PUBLICATION 61000 and CISPR 11:2009, A1:2010

**Sections Evaluated:**

Electrostatic Discharge Immunity Test  
Electric Field Radiated Immunity Test  
Magnetic Field Immunity Test  
Radiated Emissions

**PERFORMANCE MONITORED:**

Verify that the BATTERY OK light stays off and that no voltage appears at the metal plate on the end.

**Summary:**


Immunity: The Digital Voltage Indicator, Model DVI-100, met the requirements of EN 61326-1 Electrical Equipment for Measurement, Control and Laboratory Use; EMC Requirements (IEC 61326-1) for Industrial Locations using the test procedures IEC 61000-4-2, IEC 61000-4-3, and IEC 61000-4-8.

Emissions: It was found that the Model DVI-100 met the radio interference radiated emission requirements of EN 61326-1:2006 Electrical Equipment for Measurement, Control and Laboratory Use. The product represented in this test report is battery operated and has no provisions for connection, either directly or indirectly, to the AC mains power supply.

*Furthermore, the manufacturer may declare that the product is in conformity with the above sections of EN 61326-1:2013.*

Attested to:

D.L.S. Electronics

  
Brian J. Mattson  
General Manager

2-10-2017  
Date

