

**Certificate of Conformity**  
**D.L.S. Electronic Systems**

**Company:** HD Electric Company  
**Product Tested:** Rope Tester  
**Model Tested:** Rope Tester  
**Testing Date:** May 27<sup>th</sup> – 31<sup>st</sup>, 2016  
**Report Numbers:** 21932, 21933 & 21934

*A sample of the equipment as configured in the accompanying test report has been found to comply with the following standards:*

EN 61326-1:2013  
ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE;  
EMC REQUIREMENTS - Part 1: General Requirements  
(IEC 61326-1:2012) - Industrial Locations  
USING IEC PUBLICATION 61000 and CISPR 11:2009, A1:2010

**Sections Evaluated:**

Radiated Emissions Test  
Harmonic and Flicker Emissions Test  
Electrostatic Discharge Immunity Test  
Electric Field Radiated Immunity Test  
Surge Immunity Test  
Conducted Immunity Test  
Magnetic Field Immunity Test  
Voltage Dips, Interruptions & Variations Immunity Test

**PERFORMANCE MONITORED:**

Monitor the LED display for the green LED turning off, the red LED turning on or any sound from the beeper.

**Summary:**

Immunity The Rope Tester - met the requirements of EN 61326-1 Electrical Equipment for Measurement, Control and Laboratory Use; EMC Requirements (IEC 61326-1) for Industrial Locations using the test procedures IEC 61000-4-2, IEC 61000-4-3, and IEC 61000-4-8

Emissions: The Rope Tester - met the radio interference radiated emission requirements of EN 61326-1:2006 Electrical Equipment for Measurement, Control and Laboratory Use. The product represented in this test report is battery operated and has no provisions for connection, either directly or indirectly, to the AC mains power supply.

*Furthermore, the manufacturer may declare that the product is in conformity with the above sections of EN 61326-1:2013.*

**Attested to:**

D.L.S. Electronics

  
\_\_\_\_\_  
Brian J. Mattson  
General Manager

4-10-2017

\_\_\_\_\_  
Date

